



उदयपुरगढी गाउँपालिका
गाउँकार्यपालिकाको कार्यालय



पत्र संख्या:- २०८२/०८३

चलानी नं.:-

website:- udayapurgadhimun.gov.np

gmail:- udayapurgadhimun@gmail.com

दररेट पेश गर्ने सम्बन्धी सूचना ।

प्रथम पटक प्रकाशित मिति:- २०८२/०५/१७

यस कार्यालयको आ.व.२०८२/८३ मा स्वास्थ्य संस्थाहरूको लागी औषधि तथा औषधिजन्य सामग्रीहरू र अन्य सामग्रीहरू खरिद कार्य गर्नुपर्ने भएकोले, सो खरिद कार्यको लागत अनुमानका लागी लिखित रुपमा दररेट माग गर्ने सम्बन्धमा यो सूचना प्रकाशित गरिएको छ । सूचना प्रकाशन भएको मितिले ७ (सात) दिन भित्र कार्यालय समयमा तपसिल बमोजिमको शर्तहरूको अधिनमा रही तोकिएको ढाँचामा लिखित रुपमा अंक तथा अक्षरमा दररेट पेश गर्नको लागी यस कार्यालयको वेबसाईट र सूचना पाटीमा सूचना टाँस गरिएको जानकारी गराइन्छ ।

शर्तहरू:

- एकाई दरमा मु.अ.कर सहित (मु.अ.कर लाग्ने सामग्रीको हकमा) सबै भन्सार शुल्क तथा अन्य करहरू, अन्तिम गन्तव्य (उदयपुरगढी गाउँपालिकाको कार्यालय) सम्मको दुवानी खर्च समेत समावेश भएको हुनु पर्ने छ ।
- लिखित दररेटको अंक र अक्षरमा फरक परेमा अक्षरलाई मान्यता दिइनेछ ।
- फर्म / सप्लायर्सबाट पेश हुन आएको प्रति इकाई मुल्य आ.व.२०८२/८३ को आषाढ मसान्तसम्म मात्र कायम हुनेछ । बिचमा दररेट परिवर्तन गर्न पाइने छैन । एक पटक पेश गरेको दररेटमा चालु आर्थिक वर्ष भरि नै आपूर्ति कर्ताले सामग्री उपलब्ध गराउनु पर्नेछ ।
- कार्यालयले आफ्नो अनुकूल समय र परिस्थितिमा कुनै पनि बखत सामग्री खरिदका लागी खरिद आदेश जारी गर्न सक्नेछ ।

नोट: दररेट पेश गर्ने सामानको नाम, स्पेसिफिकेशन र ढाँचा यस कार्यालयको Website मा राखिएको छ ।

भक्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत
भक्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत

Udayapurgadhi R.M.

Udayapur

Dare Rate: 2082/83

S.N.	Particular	Specification	Quantity	Unit	Unit Cost (Amount in figures)	Unit Cost (Amount in words)	Remarks
1	2	3	4	5	6	7	10
1	Pt. Bed	Bed base shall be anti-corrosive and antirust treated epoxy powder coated steel frame The bed base top shall be made of epoxy powder coated 18G perforated sheet to improve ventilation The patient bed shall have a fixed height. Shall have 4 IV rod receptacles and mosquito net pole receptacles at the 4 corners. It shall come with one dual hook anti-corrosive and antirust treated epoxy powder coated or 304 grade stainless steel IV rod. Shall have provisions to fix urinary bag on both sides. All 4 legs of the bed shall be capped with heavy duty rubber footings. It shall have S.S. / epoxy powder coated bedhead & foot-end fitted with laminated panels. Both bedhead and foot-end panel shall be detachable. The height of the bedhead panel: not less than 1060mm from floor. The height of the foot-end panel: not less than 820mm from floor. Overall approximate dimension: not less than 1980mm length, 910mm width, 600mm height	1	Pcs.			
2	Hand support steel chair (5 seater)	L 1800* W 650 * H 800 mm (stainless steel structure and aluminium armrest and legrest)	1	Pcs.			
3	Bedside locker	Feet to be capped with heavy duty plastic buffers. Overall approximate size 820mm H x 400mm W x 400mm L Fitted with superimposed stainless steel top. Top to have lip or edge or retaining rail to prevent items slipping off, Finish must be smooth. With stainless steel towel rail. Lockable drawer immediately beneath the top, minimum height of drawer 18 cm. Below the drawer space open on all four sides – min 20 cm height to the cupboard to allow access from all sides. Below the open space one cupboard with metal handle/knob with reversible hinge Cabinet door so that the door direction can be adjusted to open to the right or left depending on where it is to be used.	1	Pcs.			
4	Steel table (office table)	Stainless Steel :- 6*3.5 (3 drawers)	1	Pcs.			
5	Chair hand rest	Polypropylene, L 31cm*W 10cm*H 29cm	1	Pcs.			
6	Steel Cupboard	H 182.5cm* L 96cm* B 45cm , Minimum four racks in two (Door) compartments, gauge of steel 0.8mm - 0.7mm, steel framed double hang door with lock	1	Pcs.			
7	Water tank	1000 Liter	1	Pcs.			
		Sturdy and robust table/bed for labour and delivery.					
		Approximate Dimensions: 2200L x 1000W mm.					

(Handwritten signature)



(Handwritten signature)
महोदय बहादुर राई
 प्रमुख प्रशासकीय अधिकृत

8	Delivery bed (Hydraulic)	Frame made in robust combination of tubular/oval steel tube epoxy powder coated washable paint at least 12 gauge.	1	Pcs.				
		Table top must have 3 adjustable parts:						
		Ratchet operated rising backrest, retractable foot end and a fixed centre part.						
		Two support handles shall be positioned on either side of the rising backrest.						
		Leg support stands with waterproof and washable protection.						
		Birth permitted in at least the following positions:						
		<ul style="list-style-type: none"> Classic Lithotomic. Sitting and/or semi-sitting. Knee-elbow. Vertical. Squatting. Caesarean, surgical and foetus extraction. 						
		Vertical movement control/adjustment via hydraulically operated monobloc actuator with dual-sided pedal control.						
		All movement and control via gas-spring operated by head/side levellers.						
		Quick release mechanism for back section adjustment during emergency situation for quick passage to CPR (Cardio Pulmonary Resuscitation) position.						
9	Delivery set	Head and Lower Ends:	1	Set				
		<ul style="list-style-type: none"> Easily removable (for resuscitation). Made of Stainless Steel with plastic-covered bends and grips. Removable lower end for easy access to the delivery. 						
		Shall have provision of IV pole attachment.						
		The bed shall be mounted on four anti-static castor wheels of dia. approx. 100mm, two with brakes.						
		Three separate mattresses of at least 100 mm thickness for each section fixed in position to the top of the bed by Velcro strips on the underside of each mattress section, pad covering material to be anti-bacteria, anti-static, waterproof, fire retardant and washable.						
		The Delivery bed shall have maximum weight at least 200Kg.						
		Artery forceps-16cm-1, cord cutting scissor-1, sponge holding forceps-20cm-1, Bowel S.S (small 600ml and big 750ml)-2, Kidney tray-1, Thread for cord tie-1, Wrappers (baby and mother)-4, Rubber sheet-1						
		Episiotomy Scissors-Braun-Stadler-1 (one each of 14.5cm & 22cm), Needle -2 (mayo hegar (16cm), Dissecting forceps -1 (toothed and non toothed-14cm), Stitch cutting Scissors-2 (abdominal, kelly) 16cm, Sponge holding forceps-1 (forester, straight 25cm), Small bowel-1 180ml, Kidney tray-1						
		Artery forceps-1(140mm), Toothed thumb forceps-1(155mm), Suture cutting scissors-1 (140mm), Kidney tray-1(400mm), Bowel-1(180ml)						
		Sponge holding forceps (forester, straight, serrated)-4 (20cm), Artery forceps-1 (haemostatic, rochester-pean) (20cm), Needle holder -1 (mayo-hegar) (20cm), Scissors-1 (abdominal kelly) (18cm), Suture cutting scissors (Long)-1, Dissecting forceps-1 (non-toothed pecksmith) (15cm), Sims vaginal speculum (18cm)						
10	Episiotomy set	Episiotomy Scissors-Braun-Stadler-1 (one each of 14.5cm & 22cm), Needle -2 (mayo hegar (16cm), Dissecting forceps -1 (toothed and non toothed-14cm), Stitch cutting Scissors-2 (abdominal, kelly) 16cm, Sponge holding forceps-1 (forester, straight 25cm), Small bowel-1 180ml, Kidney tray-1	1	Set				
11	Dressing set	Artery forceps-1(140mm), Toothed thumb forceps-1(155mm), Suture cutting scissors-1 (140mm), Kidney tray-1(400mm), Bowel-1(180ml)	1	Set				
12	Cervical tear repair set	Sponge holding forceps (forester, straight, serrated)-4 (20cm), Artery forceps-1 (haemostatic, rochester-pean) (20cm), Needle holder -1 (mayo-hegar) (20cm), Scissors-1 (abdominal kelly) (18cm), Suture cutting scissors (Long)-1, Dissecting forceps-1 (non-toothed pecksmith) (15cm), Sims vaginal speculum (18cm)	1	Set				

2/12/20



मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

13	Foot step (Double)	It shall be made of anti-corrosive and antirust treated epoxy powder coated steel with a tubular frame with heavy-duty washable finishes. Dimension: approximately 45 H x 45 W x 45 D cm. The foot step shall stand on all legs at the same time on a level.	1	Pcs.				
14	Mattress	The mattress negotiable with hospital beds having an overall approximate dimension of not less than 1980mm length, 910mm width. It shall have a high density approx. 40kg/m ³ PU foam mattress. The mattress shall have thickness of at least 100mm. The corners of the mattress shall be round shaped. Mattress with two sections shall be designed to bend with the positioning of the bed when the backrest and foot section of the bed are adjusted. The mattress shall be fire retardant, antibacterial treated. The weight capacity of the mattress shall be more than 100kg. Mattress Cover: The mattress cover shall come with a zipped fire retardant, antibacterial, antistatic, acid resistance, waterproof and washable vinyl or vinylized nylon cover. The mattress cover shall be designed to provide ventilating airflow over the patient's skin. The mattress cover zip shall be a heavy-duty/large toothed synthetic zipper to enable inspection of the inner foam and totally covered by a flap extending over the zip to prevent ingress of fluids into the actual mattress and allow for replacement. The colour of the mattress cover black	1	Pcs.				
15	Pillow	Pillow cotton (at least 1kg) with sealed vinyl/plastic cover, fully washable surface and completely waterproof. Size approx.: 40x60cm and approx. 15cm thick in fire retardant antibacterial anti-mite treated foam rubber. The pillow shall come with a zipped fire retardant antibacterial, antistatic, acid resistance, waterproof and washable vinyl or vinylized nylon pillow cover. Bidder shall indicate the material of cover offered.	1	Pcs.				
16	Sunmica covered wooden table	32" inch L * 24" B * 17 mm H		Pcs.				
17	Suction machine (electronic)	It shall be mounted on four robust, fully 360 degree swivelling, antistatic, non-marking grey tires castors, minimum size 75 mm with at least 2 diagonal brakes. Come with suction controller and vacuum gauge / indicator. The pump shall be oil free vacuum pump where the pumped liquid shall be sealed off from the pump. Come with overflow control valves. Bidder shall provide technical design and details of the pump with this TSF Vacuum rate shall be from 0 to not less than 640 mmHg (0.85 bars). Air flow rate shall be at least 25 l/min. The pump shall come fitted with twin unbreakable, transparent, autoclaveable polycarbonate suction bottles minimum 2 litre each. The bottles shall be incorporated with an automatic suction cut-off mechanism when they become full.	1	Pcs.				



(Signature)
भक्त गंगाधर राई
 प्रमुख प्रशासकीय अधिकृत

		The suction bottles shall come with overflow lid.				
		Noise level: not more than 55 dBA.				
		Air discharge from pump shall be filtered by a 0.3 micron bacterial hydrophobic filter.				
18	Suction tube (disposable) big and small size		1	Pcs.		
19	Oxygen cannula and Mask (big and small size)		1	Pcs.		
20	Patient examination table	<p>The Bed shall be made of a solid steel sheet and plate construction with anti-corrosive and antirust treated epoxy powder coating with upholstered top.</p> <p>All 4 legs of the bed shall be capped with heavy duty rubber footings.</p> <p>Overall size of the table shall not be less than 1830 mm L x 600mm W x 825 mm H.</p> <p>Strong Mild steel tubular construction epoxy powder coated treated. The top base of machine pressed double bent Mild steel sheet epoxy powder coated treated finish.</p> <p>The mattress shall be foldable and shall be designed to bend with the positioning of the bed when the backrest of the bed is adjusted.</p> <p>Bidder shall indicate the weight capacity and the total weight of the mattress in kilogram (kg).</p> <p>The mattress shall have mid-firmness, with foam density of approximately 0.55kg/ cubic foot, to avoid that the patient would sink down into foam with antibacterial, antistatic, acid resistance, waterproof and washable vinyl or vinylized nylon cover.</p>	1	Pcs.		
21	IV Stand	<p>The IV stand shall be made of tubular anti-corrosive and antirust treated epoxy powder coated mild steel, with a 5 pronged base fitted on mobile on swivelling castors of approx. diameter Ø50mm. The castors must be non-rusting and non-marking.</p> <p>The stand should come with stainless steel double IV hook, height adjustable from approximately 1620mm to 2340mm, with a screw knob for height adjustment.</p>	1	Pcs.		
22	Peri Light	<p>Light weight and easy to carry.</p> <p>The lamp can easily be positioned up to 40 degrees backwards.</p> <p>Power: 50 Watt or more with extra focus. Bidder to specify the power of lamp.</p> <p>Shall have on/off switch.</p> <p>Shall have facility to focus the light on specific area.</p> <p>Shall have long lifespan of lamp. Bidder to specify the lifespan of lamp.</p>	1	Pcs.		
23	Oxygen Cylinder 8F		1	Pcs.		
24	BP instrument	<p>Packed in easy carrying high quality pouch made of waterproof cloth to accommodate cuff, and inflation bulb.</p> <p>Gauge to be calibrated in 2 mm Hg units.</p> <p>Must provide blood pressure cuffs for child size and for adult size (regular).</p>	1	Pcs.		
		Dual, cup/bell and diaphragm head				

10.9



मन्त्रि बहादुर राई
प्रमुख प्रशासकीय अधिकृत

25	Stethoscope	Head and ear tube assembly to be made of non-ferrous metal, Tubes to be synthetic material and ear tubes to have shaped plastic cushion ends.	1	Pcs.			
26	Silver fetoscope (Aluminium)	Must be standard quality lightweight Aluminium. Workmanship: The surfaces of the Fetoscope shall be free from scales, burrs, pits. Edges shall be smooth rounded off and shall not be sharp.	1	Pcs.			
27	Measuring tape		1	Pcs.			
28	Trolley Medicine	It shall be constructed fully with 304 grade stainless steel sheet and tube or better. Overall size: approximately 900 H x 460 W x 760 L mm. Flat top of SS and at least 6 inch deep removable bucket at bottom. Multiple drawers (minimum 6) made of high quality materials with telescopic channels, below the platform. Minimum 6 small boxes to keep drugs at eye level. Equipped with waste bin, needle disposable container, file cassette, atraumatic corner buffers and guard rails. Shall have provision for hanging one IV fluid bottle. Equipped with lock key system. Shall be mobile on 4 x 100mm diameter (approx.) robust 360 deg. anti-rust, anti-static, noiseless, swivel castors with non-marking grey tyres and with at least 2 diagonal castors shall have brakes.	1	Pcs.			
29	Trolley Emergency (Emergency Crash Cart)	Dimensions: approx. 960mm L x 500mm W x 1545mm H. Stainless steel top and shelf & equipped with 4 - 6 removable coloured bins made of moulded plastic. Lockable storage units - 3 drawers (stainless steel or moulded plastic). Wood or wood laminate construction drawers are NOT acceptable. To be fitted with stainless steel, height adjustable, twin hook/loop, IV pole assembly. Fully, 360 deg. swivel castors/wheels, size 125mm dia with at least one castor/wheel to have locking/brake mechanism. SS shelves 14 G thick stainless steel. Top shelf to have stainless steel guard rail approx. 35mm above surface. Fitted with epoxy powder coated oxygen cylinder holder and electric lamp holder with clamp and cardiac massage board. Must be capable of carrying ECG Monitor/defibrillator and a suction apparatus.	1	Pcs.			
30	Cheatele forcep with Jar, stainless	Material: Stainless steel, autoclaveable. Conical shaped with broad base for keeping cheatele forceps of size 270 mm.	1	Pcs.			
31	Instrument Trolley	It shall be constructed fully with 304 grade stainless steel sheet and tube or better. Overall size: approximately 860 H x 460 W x 760 L mm It shall be have 2 tiers of grade 304 stainless steel shelves, top approx. at 880mm and lower shelf at 400mm. On three sides of shelves 20 mm upright lips/rail. Fourth side to have turned down edge	1	Pcs.			

Handwritten signature/initials.



Handwritten signature and official stamp of the Government of India, Ministry of Health and Family Welfare, with text in Hindi and English.

		Shall be mobile on 4 x 50mm diameter (approx.) robust 360 deg. swivel castors with non-marking grey tyres and with at least 2 diagonal castors shall have brakes					
32	Surgical Drums big size	<p>Drum must have an effective closing lid with a clip lock, a carrying handle, air vents system to allow steam to circulate freely during the sterilisation cycle.</p> <p>Vents to be manually closed after sterilisation.</p> <p>Air vent system (opening and closure mechanism) must be efficient and easy to operate.</p> <p>Lateral air vents system is preferable to top and bottom air vents.</p> <p>Material: Austenitic stainless steel, smooth surface.</p> <p>Austenitic stainless steel composition: approx. 8 to 10 % nickel, 18 to 20 % chromium.</p> <p>External diameter: approx. 340 to 350 mm.</p> <p>Height: approx. 180 to 200 mm.</p> <p>Thickness: approx. 0.6 to 0.7 mm.</p>	1	Pcs.			
33	Surgical Drums Small Size	<p>Drum must have an effective closing lid with a clip lock, a carrying handle, air vents system to allow steam to circulate freely during the sterilisation cycle.</p> <p>Vents to be manually closed after sterilisation.</p> <p>Air vent system (opening and closure mechanism) must be efficient and easy to operate.</p> <p>Lateral air vents system is preferable to top and bottom air vents.</p> <p>Material: Austenitic stainless steel, smooth surface.</p> <p>Austenitic stainless steel composition: approx. 8 to 10% nickel, 18 to 20% chromium.</p> <p>External diameter: approximately 150 to 165mm.</p> <p>Height: approximately 100 to 120mm.</p> <p>Thickness: approximately 0.6 to 0.7mm.</p>	1	Pcs.			
34	Vacuum set	<p>Hand operated device complete with floor stand jar-holding basket.</p> <p>Suction jar in unbreakable material, polycarbonate or similar –size 500ml</p> <p>1 metres of suction tubing to connect suction cup</p> <p>Pulling system; Large stainless steel or non- ferrous handle</p> <p>Silicone rubber cups requirements:</p> <ul style="list-style-type: none"> The rubber cups must be medical grade silicone rubber cups and must ensure absolute bio-compatibility and high durability Smooth external surface with small longitudinal ridge to observe any rotation inside of the cup must be lined with smooth projections 1 mm high and 4 mm in diameter, 3-5 mm distance from one another. The projections enable the air between the cup and the child's head to be evacuated and ensure that the cup sucks firmly <p>Pump requirements:</p> <ul style="list-style-type: none"> The vacuum pump must be handy and easily portable, the desired level of vacuum must be achieved quickly and the pressure maintained. The gauge must of international standard and must show accurate pressure. The pump must be highly durable and maintenance free. 	1	Pcs.			

20/09/20



महेश्वर राई
प्रमुख प्रशासकीय अधिकृत

		<ul style="list-style-type: none"> Pump to develop suction with 1 metre of connection suction tubing. Foot operated suction pump systems preferred but hand bicycle style pump acceptable. 				
		Aspirator Device Characteristic: <ul style="list-style-type: none"> Manually operated and has a holding capacity of minimal 60cc of volume. Suction capacity provides a minimum 600 mmHg (vacuum) and must maintain this vacuum for at least thirty minutes. Autoclaveable or can be disinfected by Cidex, 0.5% Chlorine or Virkon 				
		Valve Design: <ul style="list-style-type: none"> The valve liner must be removable by opening a hinged valve body. Valve must feature double locking valve allowing the vacuum to be created prior to the procedure. The valve liner must provide clear fluid path. Must be capable of being assembled, charged and ready to use 30 minutes before the procedure and not lose vacuum. 				
		Cylinder Design: <ul style="list-style-type: none"> Aspirator must feature secure stopping mechanism for plunger when pulled. Collar stop must be able to be displaced or removed for removal of plunger. Device must be engraved with manufacturer's logo and autoclave temperature. Cylinder must be marked in 10cc increments to measure evacuated uterine contents. Plunger Design: Must have ergonomic handle. 				
35	MVA Set	Reuse: <ul style="list-style-type: none"> Aspirator must have passed laboratory testing of 100 HLD or Sterilization processing cycles. Must be suitable to be reused a minimum of 25 times under normal use and maintenance. 	1	Pcs.		
		Cannulae Specifications: <ul style="list-style-type: none"> Sizes: 4, 5, 6, 7, 8, 9, 10, and 12 mm, Colour Coded 223mm long (+/- 7 mm), with a wall thickness of 0.5 to 1.3 mm depending on the size of the cannulae. Body of cannulae with one or two apertures at the tip. Must be made of Polypropylene 				
		Cannulae Design: <u>size 4 through 8:</u> <ul style="list-style-type: none"> Have two "whistle cut" type openings and do not have any sharp edges or "flash" at the openings. Openings are on opposite side of cannulae but placed such that the strength of the cannulae tip is not compromised. <u>size 9 through 12:</u> <ul style="list-style-type: none"> have a single large "scoop cut" type opening and do not have any sharp edges or "flash" <u>All:</u> <ul style="list-style-type: none"> feature dots as indicators for measuring the depth of the uterus on the cannulae. The first dot is 6 cm from the tip of the cannula; others follow at 1cm intervals. 				

35



मन्त्रि बहादुर राई
प्रमुख प्रशासकीय अधिकृत

		• Cannulae have tight, leakage-free fit with vacuum aspirator.					
36	MRP Set		1	Pcs.			
37	Wall Clock	Standard	1	Pcs.			
38	Resuscitation bag and Mask (baby size-0-1)		1	Pcs.			
39	Infant baby Warmer	Power rating: AC – 220v-230v 50Hz	1	Pcs.			
		It should have microprocessor based servo and manual control of infant ven.					
		It should have indication of LED display of infant ven in degree C or F					
		It should have heating element ceramic type (500 watts)					
		It should have heater output bar indicator					
		It should have Safety MCB switch for protection					
		Basinet side cover can be open & close without difficulty					
		It should have LED lamp should be adjustable					
		The infant's skin color should be visible clearly and light should reach directly to the infant ven canopy is tilted.					
		Skin Temp. Setting range 34 ~ 38 C					
		Display Temp. Setting range 34 ~ 40 C					
		Audible and visual alarm functions for safety					
		Power failure					
		Temperature low /high					
		Temperature sensor failure					
		It should have x-ray cassette tray.					
		The pressure dispenser mattress should be supplied					
		It should have preheating function to warm the mattress.					
		It should have Apgar & CPR timer					
		It should have Adjustable Controls for Alarm Setting by Touch screen					
		The side rails of the warmer should be foldable to care the baby and transparent.					
		It should have IV pole, made by S.S.					
		The infant radiant warmer stand should have 4 wheel of at least 2 wheel should be lockable.					
		Required accessories:					
		Power cord length 3 m. with UK Plug					
		Skin temperature probe –2 set					
Dust cover							
40	Resuscitation set	Self-inflating neonatal resuscitator bag, Hand operated, portable, reusable	1	Pcs.			
		Translucent and made from silicone rubber, capacity approx. 220 – 320ml to provide positive pressure to pre term babies, low birth weight and term babies< 5kg					
		Valve- to prevent excessive pressure, which can cause lung damage, with a pressure limiting valve so that the airway pressure does not exceed 4.5 kPa (45 cmH2 O) and can generate an airway pressure of at least 3 kPa (30 cmH2 O)					

Signature



महक बहादुर राई
प्रमुख प्रशासकीय अधिकृत

		Mask-The mask or facemask is connected to the bag and covers the mouth and nose of the newborn who is being resuscitated -Size 0 for preterm and low-weight baby, round type, diameter 35-50 mm - Size 1 for term baby, round type, diameter 50-65 mm translucent				
41	Baby blanket	double layer blanket (107 * 136 cm) thick	1	Pcs.		
42	Waste Bin- pedal operated stainless steel	Bin shall be in cylinder shape with approximately 250-350mm in diameter or square/ oblong shape with approximately 250-350mm square, height 400-550mm. It shall come with a stainless steel inner bucket (bin) with a fold-out handle. It shall have a foot operated pedal opening lid. Pedal to be minimum Ø 60mm or Ø70 mm. Lid to be in stainless steel or other non-ferrous metal. Lid or bin body must have rim cover with plastic gasket so that it does not make noise when opened/closed. Bin must have heavy base so that it will not tip over when pedal is operated and bin is empty.	1	Pcs.		
43	Bed sheet with pillow		1	Pcs.		
44	Makintosh	1 meter	1	Pcs.		
45	Bed Screen 3 fold	Three fold ward screen approx. total size 2450 w x 1650 h mm in three sections. Mild steel tubular construction with epoxy powder coated treated in three section 600mm span width at each side and 1210 mm span width in the middle) and mounted on wide spaced legs for each section, and each foot to have two swivel castors size 50mm. To be supplied with hooks, springs and heavy duty curtain, firmly attached at sides, top and bottom. Curtain must have no gaps between sections	1	Pcs.		
46	Thermometer Digital (Clinical)	Flat type, wide thermometer, safe to use, no glass, no mercury. Scale: Celsius scale. Measurement range: 32°C to 45°C Accuracy: +/- 0.1°C between 35°C to 42°C. Display: Liquid crystal display, easy to read. Shall work on a battery. Shall have a low battery indicator. Shall have facility of beep sound and switch off. Water proof for ease of cleaning.	1	Pcs.		
47	Plastic apron	Straight apron with bib, back fastening and neckband. Moisture-proof and stain resistant. Back fastening band and neckband must be strong and not detachable. Material: Opaque or translucent high quality plastic material. Blood, water and chemical resistant. Size: <ul style="list-style-type: none"> Standard adult size. Length: approximately 120cm (from top of the bib to lower edge of the apron) Width: approximately 80cm Thickness: 0.15 to 0.30mm Reusable, machine washable.	1	Pcs.		



महत्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत

		Non-sterile.					
48	Gum boot (Black)	Medium Size	1	Pair			
49	Slippers	Medium Size	1	Pair			
50	Torch light with battery	Standard	1	Pcs.			
51	Suture Set	Kidney Tray 8", Needle Holder 6", Tooth forcep 5" scissor curved 4"/ 100 mm	1	set			
52	Foot Suction	Standard	1	Pcs.			
53	Light examination, mobile (LED)	<p>Mobile examination light with sturdy construction and easily moveable.</p> <p>Shall have heavy base with 5 swivel castors, 2 with brakes. Caster must be medical chemical resistant.</p> <p>Low centre of gravity for optimal stability and reach.</p> <p>Shall have single lamp with 7 LED 12V 1W light.</p> <p>LED shall have life time more than 20,000 hours of operation.</p> <p>Field-of-view diameter, approximately. 0.15m.</p> <p>Homogeneous illumination across entire field-of-view, approx. 60.000 lux (at 0.5m).</p> <p>Colour temperature, approximately: 4500K.</p> <p>Light head mounted on spring loaded articulating arm, height approx. 1.60m.</p> <p>On/off switch incorporated in base or spring loaded articulating arm.</p>	1	Pcs.			
54	Digital pan weighing machine	<p>Large pan, approx. 450mm long and 300mm wide.</p> <p>Pan to have flat base with lips on sides only (NOT on ends).</p> <p>Lips to be approx. 80mm high</p> <p>Pan made of Acrylic, Moulded Engineering Plastic or Metal with an easy to clean surface</p> <p>Scale to weigh 0 to 6 Kg in increments of 50g</p> <p>Dial type or Danish yard-arm balance types accepted</p> <p>Shall have Tare/Zero adjustment system</p>	1	Pcs.			
55	Room thermometer	<p>Mercury free(digital) wall fixed thermometer.</p> <p>Able to read -5 to 50 degrees Celsius in large clear numbers.</p> <p>Holder for wall fixing</p>	1	Pcs.			
56	Bowel stand, Single tier	<p>Movable stand to hold a single bowl.</p> <p>Heavy carriage mounted on 5 swivel anti-static castors.</p> <p>Fixed height support column provided with ring, in which the bowl fits.</p> <p>Capacity of bowl: 3 litres. Autoclaveable</p>	1	Pcs.			
57	Kick Bucket	<p>Waste bucket with lever, on heavy mobile holder frame.</p> <p>Wide base on 5 swivel anti-static castors.</p> <p>Pivotal lever is integrated with bucket through robust hinges.</p> <p>Bucket sets and removes smoothly from holder frame.</p> <p>Holder frame is finished with continuous encircling rubber bumper.</p> <p>Materials:</p> <ul style="list-style-type: none"> Fully Stainless steel. Bumper: rubber. 	1	Pcs.			



महोदय राई
प्रमुख प्रशासकीय अधिकृत

		Dimensions, Approx. \pm 10%: <ul style="list-style-type: none"> Overall: 400 x 360mm (diameter x h) Swivel castors, diameter: 50mm Bucket capacity: 10 litres 				
58	Mayo Trolley	<p>The mayo trolley shall be constructed fully with 304 grade stainless steel or better.</p> <p>Come with a removable 304 grade stainless steel tray size approximately 560mm x 400mm x 20mm</p> <p>It shall be height adjustable to provide tray height in range minimally from 750mm to 1300mm.</p> <p>Shall come with a screw locking knob for height adjustment.</p> <p>Shall be mobile on 4 x 50mm robust fully 360° swivel castors with non-marking grey tyres and at least 2 diagonal castors shall have brakes.</p>	1	Pcs.		
59	Trolley Dressing	<p>Heavy carriage mounted on 4 swivel anti-static castors, of which two with brakes.</p> <p>Fit on both sides with push bar-handle.</p> <p>Top and bottom shelves with guard rails.</p> <p>Protective bumpers at all four corners.</p> <p>Materials:</p> <ul style="list-style-type: none"> Frame and tray: Austenitic stainless steel 18/10. <p>Dimensions:</p> <ul style="list-style-type: none"> Overall: Approximately (\pm 10%) 900 x 550 x 1000mm (l x w x h). Frame, diameter: 30mm. Thickness of shelves: Approximately 1.5mm Swivel castors approx. diameter: 100mm. Carrying capacity: Approx. 100kg. 	1	Pcs.		
60	Penguin suction	size:- pediatric	1	Pcs.		
61	KMC wrapper , 4 meter	soft fabric of at least 4m for binding the new born babies. Sample will be checked during verification.	1	Pcs.		
62	KMC Chair	<p>The height and unfolded full measurements should be approx 45 cm and approx 161cm respectively.</p> <p>The base should be approx. 74cm/45cm.</p> <p>The measurements (L/W) of back, setting base and footrest should be approx 52cm/45cm, 52cm/53cm and 56cm/52cm respectively.</p> <p>The diameter of moulds for the speakers and the head rest should be 7cm and 4cm respectively.</p> <p>The 2 built-in KMC cloths, and 3 velcro strips should be 78cm/40cm and 50cm/2.5cm.</p> <p>The side platforms for the feeding utensils should be 11.5cm/22cm and had sockets for the placement of cup.</p> <p>The area for storage of the clothes under the chair should be approx. 48cmX74cmX24cm (85258 cubic cm).</p> <p>The chair should be constructed with fabricated polished iron, good quality sponge and regzine.</p>	1	Pcs.		
63	Head Light	<p>LED head light consists of a powerful 3W LED bulb with adjustable spot size and angle of illumination.</p> <p>It must work on re-chargeable batteries as well as directly on AC mains supply using an AC adaptor/charger.</p> <p>Operating life of LED must be minimum 5,000 hrs.</p>	1	Pcs.		

5/12/23



मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

64	Ophthalmoscope Set	<p>Ophthalmoscope set shall have diagnostic head threaded on a handle.</p> <p>Head shall have wheel with lens dioptries (0 to +20 and 0 to -20), apertures small, large and semi-circle, fixation star and green filter.</p> <p>Shall have halogen bulb 2.5V with bright white light.</p> <p>Handle shall have on/off switch.</p> <p>Shall work with 2 AA-batteries (1.5V / LR6 alkaline).</p>	1	Pcs.			
65	Speculum Nasal	<p>Material: High grade fully stainless steel, corrosion resistance.</p> <p>Finish: Mirror finish.</p> <p>Autoclaveable.</p> <p>Shall provide infant / child size, 130mm long, straight, double spring.</p>	1	Pcs.			
66	Speculum Aural	<p>Material: High grade fully stainless steel, corrosion resistance.</p> <p>Finish: Mirror finish.</p> <p>Autoclaveable.</p> <p>Shall provide set of 4 different sizes of aural speculum.</p>	1	Pcs.			
67	Tongue Depressor Disposable	<p>Wooden tongue depressor, with rounded extremities.</p> <p>Size: approx. 16 x 140mm.</p> <p>Thickness: approximately 2mm.</p> <p>Single use.</p> <p>Non sterile.</p> <p>Packaging: Shall supply one box of 500 wooden tongue depressors.</p>	1	Pkt			
68	Sphygmomanometer (BP Apparatus) Digital	<p>It shall have one touch operation.</p> <p>It shall measure the systolic, diastolic and pulse rate and same shall be displayed on screen.</p> <p>Display: LCD digital display.</p> <p>Measurement method: Oscillometric type.</p> <p>Measurement range:</p> <ul style="list-style-type: none"> Pressure 0 – 290 mm Hg. Pulse : 40 – 180 beats /min. <p>Accuracy:</p> <ul style="list-style-type: none"> Pressure: + 3 or – 3 mm Hg. Pulse rate: + 5 or – 5% of reading. <p>Inflation: Automatic by electric pump.</p> <p>Deflation: Automatic pressure release valve.</p> <p>Pressure detection: Capacitative pressure sensor.</p> <p>Shall work on 4 nos. AA alkaline batteries with minimum life for 500 readings.</p> <p>It shall have memory to store last three measurements.</p> <p>It shall have auto shut off function.</p> <p>Shall come with suitable sizes of reusable arm BP cuffs of adult and paediatric.</p> <p>Shall have facility of USB port.</p>	1	Pcs.			
69	Cuff, Children and baby for Sphygmomanometer	<p>Set of Velcro cuffs in three sizes: child, baby and infant</p> <p>The cuffs are double Velcro fastening, enabling it to be adjusted to fit tightly around the arm.</p>	1	Pcs.			



(Handwritten signature)

भक्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत

	handheld	The cuff is washable, very strong and reinforced at both ends.					
70	Hammer Reflex	Hammer, reflex testing, Taylor type, regular size approximately 18cm. Solid metal handle, chrome plated, solid rubber head.	1	Pcs.			
71	Room heater (Quartz Trolley hrater)	2500 wt (5 heating element)	1	Pcs.			
72	Glucometer	The strip should be able to use capillary blood samples. Control solution/calibre stick to check the reliability of the strips to be supplied and quantity should be appropriate for 500 tests. Standard manual, carrying case, one set of standard battery to be supplied with the item.	1	Pcs.			
73	Pulse Oxymeter	It shall be portable unit. Display- LCD, backlight illuminated. Parameters and waveform displayed- SpO2, pulse rate, system status, plethysmogram, menus for user settings. SPO2 range: 70-100 %. Accuracy of SPO2: 3%. Pulse rate range must be 30-240bpm. Audio-visual alarms: High/low SpO2 and pulse rate, sensor off, sensor failure, low battery. Shall have alarm override facility. It must be suitable to operate in the presence of potentially flammable anaesthetic gases, and it shall not cause fire or explosion during operations. RS 232C interface for data communication. Shall have integrated printer. Inbuilt rechargeable battery and shall have battery back-up for at least 4 hours. Battery charger along with AC adaptor to be provided if integrated charger is not there.	1	Pcs.			
74	Bata (red)	10 Liter	1	Pcs.			
75	Bata (green)	10 Liter	1	Pcs.			
76	Bata (blue)	10 Liter	1	Pcs.			
77	Plastic jug	1 Liter	1	Pcs.			
		Electric steam steriliser, approximately 15L+/-10% chamber capacity with 25L+/-10% gross capacity, with drums sterilising. Bidder shall indicate here the dimensions of the chamber as well as the overall steriliser in the format of Height x Diameter in mm and chamber volume in litre. Metal vessel with high-pressure seal suitable for carrying out sterilisation using superheated steam. Operating temperature 121 °C – 134 °C pressure 1.1 to 2.2 kg/cm2 of steam pressure, and shall be used with distilled water. The cover and bottom of the sterilizer shall be made of heavy cast aluminium The container insert shall be made of Aluminium alloy. Seal: rubber gasket sealing or metal-to-metal sealing. Safety clamping locks: retaining bayonet clamp and (6) Bakelite wing nuts to prevent removal of the cover while the sterilizer is under pressure. Removable cover equipped with: Bakelite top handle.					

8/11/20



मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

78	Autoclave Electronic	<p>Dial gauge: geared steam pressure gauge, graduated in kg / cm³, PSI and degrees Fahrenheit, with colour-coded gauge showing sterilizing zone (green) and caution zone (red).</p> <p>Control valve and flexible metal exhaust tube.</p> <p>Excess pressure relief valve and over-pressure rubber plug.</p> <p>Chamber equipped with:</p> <p>Aluminium container: plain basket with handles. approx. 285 mm diameter x 215 mm height</p> <p>Aluminium inner container rack.</p> <p>Heating element.</p> <p>Stainless steel support stand to protect the heating element.</p> <p>Scored water level mark inside chamber.</p> <p>Power supply characteristics: 220-240 V, 50 / 60 Hz, approximately 5 Amp 1000 Watts heating element.</p> <p>On-Off toggle switch. Thermostatic control system: Thermostat heat control knob for thermostat (setting 0 to 8) and pilot light (red).</p> <p>Power supply cord: 3-wire grounded cord with plug (type B). at least 3 meters long</p>	1	Pcs.			
79	Bucket with cover	50 liter	1	Pcs.			
80	Bucket	10 liter	1	Pcs.			
81	Bed pan	<p>Female form, Stainless steel.</p> <p>Standard hospital bedpan, approximately 50mm deep, minimum 20mm lip.</p> <p>Single piece, spun, SS.</p> <p>Autoclaveable.</p>	1	Pcs.			
82	Stool Revolving on castor	<p>Revolving stool, height adjustable with pre-assembled gas spring.</p> <p>Heavy carriage mounted on 5 swivel anti-static castors.</p> <p>Height adjusting lever is conveniently placed under seat.</p> <p>Seat fit with thick upholstery.</p>	1	Pcs.			
83	Portable standing spot light	<p>Portable with sturdy construction.</p> <p>Halogen bulb: 12V/35-40W long life or LED light, approx. 5" dia with reflector.</p> <p>Homogeneous illumination across entire field of view 6000-10000 lux or higher.</p> <p>It shall have light stand supported by antistatic castor wheel base with minimum for legs.</p> <p>It shall have flexible arm.</p> <p>It shall be cool touch.</p> <p>It shall have integrated transformer with ON/OFF switch.</p>	1	Pcs.			
84	IUCD Insertion And Removal Kit	<p>Gallis pot; Galli pot-4 onz, 1.5" height, SS</p> <p>Kidney Tray, 10"</p> <p>Forceps-artery, Pean, straight, 8 1/2"</p> <p>Volselum, 9.5", curved</p> <p>Sponge holder, 9.5", straight</p> <p>Mayo dissecting scissors s/s curved, 7"</p> <p>Uterine sound, 12.5", metallic</p> <p>Bi-valve Vaginal speculum medium</p>	1 1 1 1 1 1 1 1	Kit			

21/9/20



महत्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत

		Removal				
		Gallis pot; Galli pot-4 onz, 1.5" height, SS	1			
		Kidney Tray, 10"	1			
		Forceps-artery, Pean, straight, 8 1/2"	1			
		Volselum, 9.5", curved	1			
		Sponge holder, 9.5", straight	1			
		Bi-valve Vaginal speculum medium	1			
85	Implant Insertion And Removal Kit	Gallis pot; Galli pot-4 onz, 1.5" height, SS	2	Kit		
		Dissecting forceps, 6", straight, (Non-toothed)	1			
		Sponge holder, 9.5", straight	1			
		Implant trochars with canula	1			
		Syringe, disposable 5 ml with 2" needle	1			
		Artery Forceps curved 14 cm	1			
		Removal				
		Gallis pot; Galli pot-4 onz, 1.5" height, SS	1			
		Dissecting forceps, 6", straight, (Non-toothed)	1			
		Sponge holder, 9.5", straight	1			
		Surgical handle no. 3	1			
		Mosquito forceps, curved 5"	1			
		Ringed Forceps for NSV (3.5 mm)	1			
		Syringe, disposable 5 ml with 2" needle	1			
		Artery Forceps curved 14 cm	1			
86	Height board (2 in 1) (child adult height)		1	Pcs.		
87	Tooth Extraction Set		1	Set.		
88	Open medicine Rack	Four shelves, size 2130 x 910 x 380 mm	1	Pcs.		

Name of Firm

Stamp of Firm

[Handwritten signature]



[Handwritten signature]
मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

Udayapurgadhi R.M.

Udayapur

Dare Rate- 2082 /83

S.N.	Particular	Quantity	Unit	Unit Cost (Amount in figures)	Unit Cost (Amount in words)	Remarks
1	2	3	4	5	6	9
1	Needle destroyer (electric)	230 V, 50Hz, AC / 2A fuse, ABS plastic	1			
2	Chlorhexidine gluconate solution 1000ml	Bott.	1			
3	Wrapper	1 meter	1			
4	Chlorhexidine + Cetrimide 1000ml external liquid	Bott.	1			
5	Gown reusable	medium (cotton) green	1			
6	Door mat	38 x 58 cm (heavy)	1			
7	Dettol 60 ml	Bott.	1			
8	Dustbin	50 L, green,blue& red colour	1			
9	Surgical gloves (6.5 N. disposable)	Pair	1			
10-	Surgical gloves (7 N. disposable)	Pair	1			
11	Examination (Loose) gloves	Pcs.	1			
12	Hand sanitizer 5 L	Jar.	1			
13	Hand sanitizer 500 ml spray	Bott.	1			
14	Plastic apron	Size: medium	1			
15	Utility gloves	Size: Large	1			
16	Head cap(100 pcs)	Thickness: 30-40 Gsm.	1			
17	Dettol soap 75 gm	Pcs.	1			
18	soap case	Pcs.	1			
19	N-95 Mask	Pcs.	1			
20	Eye towel (green), cotton	36 x 36 - 1" hole	1			
21	Paper towel	Pcs.	1			
22	Towel cotton	380 Gsm, 2ply	1			
23	Bed bath towel-chlorhexidine impregnated (10 wet towels)	300mm x 240 mm	1			
24	Glycerin 250gm	Bott.	1			
25	Bandage cutting scissor (8 inch)	8"	1			
26	Toilet brush	diameter: 3" & Handle Lenth: 14	1			
27	Cleaning mop with handle	Aluminium handle with 2 cotton mops heads-12py (29cm)	1			
28	Toilet cleansing liquid	1 liter	1			
29	phenol black 5 L	1 liter	1			
30	Phenol	1 liter	1			
31	Bata (plastic)	diameter: 12"	1			
32	Plastic Jug	1 liter	1			
33	Electric Jug	2 liter	1			
34	Clean delivery set (Sutkeri samagri)	Pcs.	1			
35	Green cloth	1 meter	1			
36	Waste Segregation Compartment - 3 Bins	60 Ltrs (utility Compartment)	1			

[Handwritten signature]



[Handwritten signature]
महोदय महोदय
प्रमुख प्रशासकीय अधिकृत

37	Surgical Mask	3ply	1			
38	Tissue Paper	Roll	1			
39	Wheel Powder (Washing)	1 Kg	1			
40	Delivery Pad	Pkt.	1			
41	Cotton Swab	Pkt.	1			

Name of Firm

Stamp of Firm

(Handwritten signature)



(Handwritten signature in red ink)

मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

Udayapurgadhi R.M.

Udayapur

Dare Rate- 2082 /83

S.N.	Particular	Quantity	Unit	Unit Cost (Amount in figures)	Unit Cost (Amount in words)	Remarks
1	2	3	4	5	6	9
1	Mahanarayan Tel 60 ml	1	pcs.			
2	Mahayograj gugugle 25 gm	1	Phyl			
3	Hingwastak Churna 50 gm	1	Phyl			
4	Nimbadi Churna 50 gm	1	Phyl			
5	Sitopaladi Churna 50 gm	1	Phyl			
6	Chandraprabhavati 25 gm	1	Phyl			
7	Mahashankhavati 25 gm	1	Phyl			
8	Tankan bhasam 10 gm	1	Phyl			
9	Gandhak Rasayan 10 gm	1	Phyl			
10	Balamrita 110 ml	1	Pcs.			
11	Mrigamdasab 15 ml	1	Pcs.			
12	Satawari Churna 100 gm	1	phyl			
13	Jwano Churna 100 gm	1	phyl			
14	Dasmul Tel 60 ml	1	Pcs.			
15	Amalaki Churna 50 gm	1	Phyl			
16	Khadiradivati 10 gm	1	Phyl			
17	shadbindu Tel 15 ml	1	pcs.			
18	Avipattikar Churna 100 gm	1	Phyl			
19	Trifala Churna 50 gm	1	Phyl			
20	Ashwogandha Churna 100 gm	1	Phyl			
21	Pushyanug churna 50 gm	1	Phyl			
22	Mahasudarshan bati 25 gm	1	Phyl			
23	Bilwadi Churna 50 gm	1	Phyl			
24	Kaishor gugugle 25 gm	1	Phyl			
25	Sarpgandhavati 25 gm	1	Phyl			
26	Arjuna Churna 100 gm	1	Phyl			
27	Guduchi Churna 100gm	1	Phyl			
28	Jyatyadi Tel 50 ml	1	Phyl			
29	Yograj gugugle 25 gm	1	Phyl			
30	Neeri	1	Tab.			

Name of Firm

Stamp of Firm

(Handwritten signature)



(Handwritten signature)
मन्त बहादुर राई
 प्रमुख प्रशासकीय अधिकृत

Udayapurgadhi R.M.

Udayapur

Dare Rate- 2082 /83

S.N.	Particular	Quantity	Unit	Unit Cost (Amount in figures)	Unit Cost (Amount in words)	Remarks
1	2	3	4			9
1	Nebulizer Solution-Terbutaline 10mg ,10ml	1	Bott.			
2	Nebulizer Solution- Ipratropium 250mcg , 15ml	1	Bott.			
3	Nebulizer Solution- Salbutamol 5mg , 15ml	1	Bott.			
4	Inhaler-Salbutamol 200mcg	1	Pcs			
5	Rotacap-Salbutamol 200mcg	1	Phyl			
6	Rotacap- Beclometasone 200mcg	1	Phyl			
7	Inhaler-Beclometasone 200 mcg	1	Pcs			
8	Rota inhaler (device)	1	Pcs			
9	Rotacap-Budesonide 200mcg	1	Phyl			
10	Inhaler- Budesonide 200mcg	1	Pcs			
11	Salbutamol Respules 2.5mg (5x2.5ml)	1	Pkt			
12	Cefixime 200mg	1	Tab.			
13	hydrocortisone 100mg inj.	1	Vail			
14	Cefpodoxime 200mg	1	Tab.			
15	Indomethacin 25mg	1	Tab.			
16	Alprazolam 0.25mg	1	Tab.			
17	Tinidazole 500mg	1	Tab.			
18	Amlodipine 10mg	1	Tab.			
19	Frusemide 10mg/ml,2ml ,inj	1	Amp			
20	Secnidazole 1g	1	Tab.			
21	Terbinafine 250mg	1	Tab.			
22	Thiamine 100mg	1	Tab.			
23	Vitamin B. Coplex	1	Cap			
24	Cynocobalamine 1500mcg	1	Tab.			
25	Pantoprazole 40mg	1	Tab.			
26	Diclofenac 3ml, inj	1	Amp			
27	Diclofenac 3% gel 30g ,cream	1	Tube			
28	Multivitamin 3ml, injection	1	Amp			
29	Diclofenac 50gm Spray	1	Phyl			
30	Paroxicam 20mg	1	Tab.			
31	Fexofenadine 120mg	1	Tab.			
32	Fluoxetine 20mg	1	Tab.			
33	Calcium Carbonate + Vitamin D3	1	Tab.			
34	Dulcolax 5mg	1	Tab.			
35	Lactulose 100ml solution	1	Bott.			
36	Itopride 50mg	1	Tab.			
37	Domperidone 10 mg	1	Tab.			
38	Naproxen 250mg	1	Tab.			
39	Chlorhexidine 0.2% mouth wash solution 100ml	1	Bott.			
40	Crepe bandage 10cm x 4 Mtrs	1	Pcs.			



(Signature)
मन्त बहादुर राई
 प्रमुख प्रशासकीय अधिकृत

41	Terbinafine 1%, cream 15gm	1	Tube			
42	Povidone Iodine Gargle 100ml	1	Bott.			
43	Vitamin B12 Inj.	1	Amp			
44	Pantoprazole 40mg inj.	1	Vail			

Name of Firm

Stamp of Firm

Handwritten signature



Handwritten signature
गण्डकी प्रदेश सरकार
प्रमुख प्रशासकीय अधिकृत

Udayapurgadhi R.M.

Udayapur

Dare Rate- 2081 /082

S.N.	Particular	Quantity	Unit	Unit Cost (Amount in figures)	Unit Cost (Amount in words)	Remarks
1	2	3	4	5	6	9
1	Acetylsalicylic acid (Aspirin) 75mg	1	Tab.			
2	Adrenaline inj.(1:1000) 1mg in 1ml	1	Amp.			
3	Albendazole 400mg	1	Tab.			
4	Albendazole 200mg/5ml, 10ml suspension	1	Bottle			
5	Aluminium hydroxide 250 mg + Mag.hydroxide 250 mg	1	Tab.			
6	Amitriptyline 10 mg	1	Tab.			
7	Amlodipine 5mg	1	Tab.			
8	Amoxycillin 500 mg	1	Cap.			
9	Amoxycillin 125 mg DT.	1	Tab.			
10	Amoxycillin 250 mg DT.	1	Tab.			
11	Amoxycillin 125mg/5ml Dry suspension, 60ml	1	Bottle			
12	Ampicillin 500mg	1	Vial			
13	Ampicillin 250mg	1	Vial			
14	Atropine inj 0.6 mg/ml, 1ml	1	Amp.			
15	Azithromycin 500mg	1	Tab.			
16	Azithromycin 250mg	1	Tab.			
17	Calamine lotion 15% w/v, 30ml	1	Bottle			
18	Calcium gluconate 1mg/10ml	1	Amp.			
19	carbamazepine 200mg	1	Tab.			
20	carbamazepine 100mg	1	Tab.			
21	Cetirizine 10mg	1	Tab.			
22	Cetirizine 5mg/5ml, syrup, 30ml	1	Bottle			
23	Charcoal Activated powder 10 gm	1	sachet			
24	Chlorohexidine Gel 4% (Navi malaham)	1	Tube			
25	Chlorohexidine solution, 0.2% oral 50ml	1	Bottle			
26	Chloroquine 150mg	1	Tab.			
27	Ciprofloxacin 250 mg	1	Tab.			
28	Ciprofloxacin 500mg	1	Tab.			
29	Ciprofloxacin eye/ear drops 5ml 0.3%w/v	1	Vial			
30	Ciprofloxacin eye ointment 0.3%, 15 gm	1	Tube			
31	Clotrimazole Cream 25 gm 1% w/v	1	Tube			
32	Clotrimazole 100mg, vaginal Tablets	1	Tab.			
33	Clotrimazole 1% w/v, mouth paint, 25ml	1	Bottle			
34	Clove oil liquid 5 ml	1	Vail			
35	Cloxacilin 250mg	1	Cap.			
36	Cloxacilin 500mg	1	Cap.			
37	Cotrimoxazole 480 mg	1	Tab.			
38	Cotrimoxazole 960 mg	1	Tab.			
39	Cotrimoxazole 200mg +40mg/5ml, 50ml	1	Bottle			
40	Calcium 500mg	1	Tab.			

५४९



[Handwritten signature]

मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

41	Dextrose 25%,50ml	1	Bottle			
42	Diazepam 5mg	1	Tab.			
43	Diazepam 5mg/ml, 2ml	1	Amp.			
44	Diclofenac sodium 25mg/ml, 3ml	1	Amp.			
45	Diclofenac sodium 50mg	1	Tab.			
46	Doxycycline 100mg	1	Cap.			
47	Ferrous sulphate 60 mg elemental iron + Folic acid 400mcg	1	Tab.			
48	Fluconazole 150mg	1	Cap.			
49	Fluoxetine 20mg	1	Cap.			
50	Folic acid 5mg	1	Tab.			
51	Frusemide 40 mg	1	Tab.			
52	Gentamycin 40mg/ml, 2ml	1	Vial			
53	Gention Violet 1% solu 10 ml	1	Vial			
54	Glimepiride 2mg	1	Tab.			
55	Glimepiride 1mg	1	Tab.			
56	Hyoscine butylbromide 10 mg	1	Tab.			
57	Ibuprofen 400mg	1	Tab.			
58	Ibuprofen 200mg	1	Tab.			
59	Ibuprofen 100mg/5ml, syrup 100ml	1	Bottle			
60	Lignocaine injection 2%, 30ml	1	Vail			
61	Losartan 25mg	1	Tab.			
62	Losartan 50mg	1	Tab.			
63	Magnesium sulphate 500mg in 2ml	1	Amp.			
64	Metformin 500mg	1	Tab.			
65	Metformin 1000mg	1	Tab.			
66	Metoclopramide 10 mg	1	Tab.			
67	Metoclopramide inj. 5mg/ml,2ml	1	Amp.			
68	Metronidazole 400 mg	1	Tab.			
69	Metronidazole 200 mg	1	Tab.			
70	Metronidazole Susp. 60 ml 200 mg/5ml	1	Bottle			
71	Mifepristone 200mg cambipack (1tab mifepristone200mg and 1tab.misoprostol200mcg)	1	Tab.			
72	Misoprostol 200mcg	1	Tab.			
73	Neomycine skin ointment 15gm.	1	Tube.			
74	Nifedipine 10mg	1	Tab.			
75	Nitrofurantion 100mg	1	Tab.			
76	Normal saline 0.9%,500ml	1	Bottle			
77	ORS WHO Formula /1liter	1	sachet			
78	Oxymetazoline nasal drops 0.025%,10ml	1	Vial			
79	Oxymetazoline nasal drops 0.05%,10ml	1	Vial			
80	Oxytoxin10 IU/ml, 1ml	1	Amp			
81	Paracetamol 500mg	1	Tab.			
82	Paracetamol 125mg/5ml, 60ml	1	Bottle			
83	Permethrin 1%, lotion, 60ml	1	Bottle			
84	Permethrin 5% w/v, cream,30gm	1	Tube			
85	Pheniramine maleate 2ml 22.75 mg/ml	1	Amp.			

5/5/20



मन्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत

86	Povidone 450ml 5% w/v	1	Bottle			
87	Pralidoxim 500mg, 20ml	1	Vial			
88	Prednisolone 5mg	1	Tab.			
89	Prednisolone 10mg	1	Tab.			
90	Prednisolone 20mg	1	Tab.			
91	Pyridoxine 10mg	1	Tab.			
92	Ranitidine 150 mg	1	Tab.			
93	Ranitidine 25mg/ml, 2ml	1	Amp.			
94	Ringer,sLactate (RL) 500ml		Bottle			
95	Risperidone 1mg	1	Tab.			
96	Risperidone 2mg	1	Tab.			
97	Salbutamol solution for nebulisation	1	Bottle			
98	Salbutamol 2mg/ml, oral liquid, 100ml	1	Bottle			
99	Salbutamol 100mcg MDI	1	Pcs			
100	Silver Sulphadiazine Cream 25 gm 1% w/v	1	Tube			
101	Sodium valporate 200mg	1	Tab.			
102	Tetracycline 1% eye ointment, 5gm	1	Tube			
103	Tinidazole 500mg	1	Tab.			
104	Vitamin B Complex	1	Tab.			
105	Vitamin K1 10mg/ml, 1ml	1	Amp.			
106	Zinc 20 mg	1	Tab.			
107	Absorbent cotton 400gm	1	Roll			
108	Adhesive Tape 4"	1	Roll			
109	Bandage 18m*90cm	1	Than			
110	Catgut 2.0	1	Pkt.			
111	Chlorine powder 0.5% 200gm	1	Pkt.			
112	DisposableSyringe 5 ml	1	Pcs			
113	DisposableSyringe 3ml	1	Pcs			
114	Gauze 18m*90cm	1	Than			
115	IV Canulla 18G	1	Pcs			
116	IV Canulla 24 G	1	Pcs			
117	IV Canulla 22G	1	Pcs			
118	IV set	1	pcs			
119	Phenol 500ml	1	Bottle			
120	Rectified Spirit 500 ml	1	Bottle			
121	Rubber Catheter Diff. size	1	pcs			
122	Silk Thread	1	Roll			
123	Surgical blade	1	Pcs			
124	Surgical Gloves Diff. Size	1	Pairs			
125	Suture needle	1	Pkt			
126	Nylon thread	1	Pkt			
127	Xylocaine gelly	1	Tube			
128	Acyclovir ointment 15gm	1	Tube			
129	Easy Fix Tape	1	Pcs			
130	Cotton \Swab	1	Pkt			
131	Emergency Contraceptive Pills	1	Tab.			

8/9/



मन्त्र बहादुर राई
प्रमुख प्रशासकीय अधिकृत

132	Non-absorbable Suture 2-0	1	Pkt.			
133	Sodium Chloride 0.9%, 100ml	1	Bottle			

Name of Firm

Stamp of Firm



[Handwritten signature in red ink]
मन्त बहादुर राई
प्रमुख प्रशासकीय अधिकृत